

ZIBO EASTMED HEALTHCARE PRODUCTS CO.,LTD

ZIBO CITY , SHANDONG PROVINCE, CHINA

TEST REPORT

Product Name:	Sterile latex surgical gloves with powder for single use only	Date tested	23/05/2019
Presentation:	Individually packed, 50 pairs/box and 500 pairs/carton	Quantity:	60000 pairs
Size:	6.5	Sample quantity:	300 pairs
Mfg. date:	20190516	Batch No.:	20190516
Exp. date:	20240515	Inspection standard:	S-2

№	TESTS	STANDART		TEST DATA			RESULT		
		Size	Specifications	Size	Specifications				
1	Width (mm)	7.0	89±5	7.0	-			-	
		7.5	95±5	7.5	-			-	
		8.0	102±5	8.0	103			PASSED	
2	Length (mm)	7.0	Min=275; Max=280	7.0	-			-	
		7.5	Min=280; Max=289	7.5	-			-	
		8.0	Min=280; Max=290	8.0	286			PASSED	
3	Thickness (mm)	Cuff	0.10-0.30	Size	Cuff	Palm	Finger	-	
				7.0	-	-	-		
		Palm	0.10-0.30	7.5	-	-	-		
		Finger	0.10-0.30	8.0	0.20	0.21	0.25	PASSED	
4	Weight (g)	7.0	8.5±0.2	7.0	-			-	
		7.5	9.0±0.2	7.5	-			-	
		8.0	9.5±0.2	8.0	9.5			PASSED	
5	Tensile ability	Before aging	Strength	≥12.5N	Size	Before aging	After aging	-	
					7.0	-	-		
					7.5	-	-		
		After aging	Elongation	≥700%	7.5	-	-	-	
						8.0	(S) 15.6	(S) 11.4	PASSED
							(E) 860	(E) 655	PASSED
6	Water tightness	Without water leak, qualified		7.0	-			-	
				7.5	-			-	
				8.0	1499			PASSED	
7	Protein content	<180ug/g		7.0	-			-	
				7.5	-			-	
				8.0	89.5			PASSED	
8	Sterilization (by E.T.O)	Sterile		7.0	-			-	
				7.5	-			-	
				8.0	Sterile			PASSED	
9	Out viewing	Without defects such as surface without damage, fracture and foreign body, no noticeable crease. Packaging meets relevant requirements		7.0	-			-	
				7.5	-			-	
				8.0	Comply			PASSED	

Checked By:
NILI

Audited By:
WALANCE LI

Approved By:
WALANCE LI

