

# ZIBO EASTMED HEALTHCARE PRODUCTS CO.,LTD.

ZIBO CITY, SHANDONG PROVINCE, CHINA

## TEST REPORT

Product Name:	Latex surgical gloves, with powder sterile for single use only (	Date tested	26/05/2019
Presentation:	Individually packed, 50 pairs/box and 500 pairs/carton	Quantity:	4//0000 pairs
Size:	7.5	Sample quantity:	150pairs
Mfg. date:	20190516	Batch No.:	20190516
Exp. date:	20240515	Inspection standard:	S-2

No	TESTS	STANDART		TEST DATA			RESULT	
		Size	Specifications	Size	Specifications			
1	Width (mm)	7.0	89±5	7.0	-		-	
		7.5	95±5	7.5	94		PASSED	
		8.0	102±5	8.0	-		-	
2	Length (mm)	7.0	Min=275; Max=280	7.0	-		-	
		7.5	Min=280; Max=289	7.5	284		PASSED	
		8.0	Min=280; Max=290	8.0	-		-	
3	Thickness (mm)	Cuff	0.10-0.30	Size	Cuff	Palm	Finger	PASSED
				7.0	-	-	-	
		Palm	0.10-0.30	7.5	0.21	0.22	0.26	
		Finger	0.10-0.30	8.0	-	-	-	
4	Weight (g)	7.0	8.5±0.2	7.0	-		-	
		7.5	9.0±0.2	7.5	9.1		PASSED	
		8.0	9.5±0.2	8.0	-		-	
5	Tensile ability	Before aging	Strength	≥12.5N	Size	Before aging	After aging	PASSED
				7.0	-	-		
			Elongation	≥700%	7.5	(S) 15.5 (E) 870	(S) 11.3 (E) 660	
		After aging	Strength	≥9.5N	8.0	-	-	
				Elongation	≥550%	8.0	-	-
6	Water tightness	Without water leak, qualified	7.0	-	-	-	-	
			7.5	1498	-	-	PASSED	
			8.0	-	-	-	-	
7	Protein content	<180ug/g	7.0	-	-	-	-	
			7.5	89	-	-	PASSED	
			8.0	-	-	-	-	
8	Sterilization (by E.T.O)	Sterile	7.0	-	-	-	-	
			7.5	Sterile	-	-	PASSED	
			8.0	-	-	-	-	
9	Out viewing	Without defects such as surface without damage, fracture and foreign body, no noticeable crease. Packaging meets relevant requirements.	7.0	-	-	-	-	
			7.5	Comply	-	-	PASSED	
			8.0	-	-	-	-	

Checked By:  
NILI



Audited By:  
WALANCE LI



Approved By:  
WALANCE LI

