

ZIBO EASTMED HEALTHCARE PRODUCTS CO.,LTD.

ZIBO CITY, SHANDONG PROVINCE, CHINA

TEST REPORT

Product Name:	Sterile latex surgical gloves with powder for single use only	Date tested	23/05/2019
Presentation:	Individually packed, 50 pairs/box and 500 pairs/carton	Quantity:	60000 pairs
Size:	7.0	Sample quantity:	300 pairs
Mfg. date:	20190516	Batch No.:	20190516
Exp. date:	20240515	Inspection standard:	S-2

No	TESTS	STANDART		TEST DATA				RESULT
		Size	Specifications	Size	Specifications			
1	Width (mm)	7.0	89±5	7.0	90			PASSED
		7.5	95±5	7.5	-			-
		8.0	102±5	8.0	-			-
2	Length (mm)	7.0	Min=275; Max=280	7.0	278			PASSED
		7.5	Min=280; Max=289	7.5	-			-
		8.0	Min=280; Max=290	8.0	-			-
3	Thickness (mm)	Cuff	0.10-0.30	Size	Cuff	Palm	Finger	PASSED
				7.0	0.21	0.22	0.25	
		Palm	0.10-0.30	7.5	-	-	-	
		Finger	0.10-0.30	8.0	-	-	-	-
4	Weight (g)	7.0	8.5±0.2	7.0	8.6			PASSED
		7.5	9.0±0.2	7.5	-			-
		8.0	9.5±0.2	8.0	-			-
5	Tensile ability	Before aging	Strength	≥12.5N	Size	Before aging	After aging	PASSED
					7.0	(S) 15	(S) 11	
				(E) 850	(E) 650	PASSED		
		After aging	Elongation	≥700%	7.5	-	-	-
						-	-	-
			Strength	≥9.5N	8.0	-	-	-
	Elongation	≥550%		-	-	-		
6	Water tightness	Without water leak, qualified		7.0	299			PASSED
				7.5	-			-
				8.0	-			-
7	Protein content	<180ug/g		7.0	90			PASSED
				7.5	-			-
				8.0	-			-
8	Sterilization (by E.T.O)	Sterile		7.0	Sterile			PASSED
				7.5	-			-
				8.0	-			-
9	Out viewing	Without defects such as surface without damage, fracture and foreign body, no noticeable crease. Packaging meets relevant requirements.		7.0	Comply			PASSED
				7.5	-			-
				8.0	-			-

Checked By: NILI



Audited By: WALANCE LI



Approved By: WALANCE LI


